

progressive diagnostics



STATEMENT OF ATTAINMENT

A Statement of Attainment is issued by a Registered Training Organisation
when an individual has completed one or more accredited units

This is a statement that

Lauren Filippe

has attained

HLTPAT005 Collect specimens for drugs of abuse testing

DARREN LEE
Director



NATIONALLY RECOGNISED
TRAINING

Issued on:
4th June 2021

Document Number:
B240121LF



Breathalyser Calibration

Certificate No: CAL013065

This certifies that the instrument listed below has been serviced and calibrated by Breathalyser Sales & Service Pty Ltd in accordance with the manufacturer's service and calibration procedures and in compliance with Australian Standard AS3547 maximum permissible error stipulations.

Customer: Dab Services

Serial No: A432369

Model: AlcoQuant 6020+ PS v121

Service Report Details

Date: 3/09/2021

Next Calibration Due: 4/03/2022

Technician: Jeans, Aimee

Batteries Replaced: New Unit

Calibration Type: Dry Gas

Calibration Gas Certificate / Batch # QCSPC024789/LGP01503-1

Calibration Gas Value: 214.000 PPM 0.082 BrAC

Final Device Reading: 0.082 BrAC

Faults/Repairs/Comments:

*NOTE: BrAC - Breath Alcohol Concentration expressed in g/210L (gram of alcohol in 210 litres of breath)
BAC - Blood Alcohol Concentration expressed in g/100ml (grams of alcohol in 100 millilitres of blood)*

We recommend that this Certificate be kept and filed for future reference.

Sydney - Head Office

128 O'Riordan Street,
Mascot NSW 2020.
Tel: 02 8338 1555
Australia Wide: 1300 999 200
nsw@breathalyser.com.au

Melbourne

29 Catalina Drive,
Tullamarine VIC 3043
Tel: 03 8609 9811
vic@breathalyser.com.au

Perth

486 Great Eastern Hwy
Ascot WA 6104
Tel: 08 9479 3866
wa@breathalyser.com.au

Brisbane

11/ 17-19 University Drive,
Meadowbrook QLD 4131
Tel: 07 3200 4944
qld@breathalyser.com.au